

# C2089-0511

Determination of Local Anesthetics by HPLC-UV together with ChraSil B100 C8, 150mm x 4.6mm, 5µm

<p>Column : ChraSil B100 C8, 150mm x 4.6mm, 5µm</p> <p>Part No. : <a href="#">C1515-B5032</a></p> <p>Mobile Phase : CH<sub>3</sub>OH:5 mM NH<sub>4</sub>HCO<sub>3</sub>, pH 10 - 65:35</p> <p>Flow Rate : 1.0 mL/min</p> <p>Temperature : Ambient</p> <p>Detection : UV, 210 nm</p>	<p><b>Peak Identification</b></p> <ol style="list-style-type: none"><li>1. Benzocaine</li><li>2. Procaine</li><li>3. Chlorocaine</li><li>4. Mepivacaine</li><li>5. 4-Hydroxyprocaine</li><li>6. Cocaine</li><li>7. Lidocaine</li><li>8. Ropivacaine</li></ol>
<p><a href="#">image_1627385736994.png</a></p> <p>ChraNotes</p>	

Revision #1

Created 27 July 2021 11:33:30 by Admin

Updated 27 July 2021 11:37:01 by Admin