


Product Specification Sheet

by ChraWork a Gimium™ Ecosystem

2002000164

LC COLUMN QUALITY CONTROL PERFORMANCE		
COLUMN INFORMATION SERIAL NUMBER : 2002000164 PART NUMBER : C1505-B5034 COLUMN TYPE : ChraSil B100 C18, 250 x 4.6 mm, 5.0um LOT : KL0051		TEST CONDITIONS MOBILE PHASE : 0.1% TFA / 99.9% H ₂ O COLUMN FLOW RATE : 1.0 mL/min COLUMN FLOW RATE : 1.0 mL/min LINEAR VELOCITY : 0.1 cm/s TEMPERATURE : 30°C INJECTION VOLUME : 10 µL
TEST VALUES THEORETICAL PLATES : 23067 SELECTIVITY : 3.27 USP TAILING FACTOR : 1.01 (@ 5% Peak Height)	SPECIFICATIONS MIN : 92268 RANGE : 3.02 - 3.62 RANGE : 0.98 - 1.25 k' : 2.21	Peak 1. 2. 3. 4.
 Image not found or type unknown		
MATERIAL SAFETY DATA SHEETS ARE AVAILABLE		
ChraReport at https://chracert.com/2002000164		

Revision #3

Created 10 August 2021 12:14:39 by Admin

Updated 6 October 2021 04:26:22 by Admin