


Product Specification Sheet

by ChraWork a Gimium™ Ecosystem

2002000204

| LC COLUMN QUALITY CONTROL PERFORMANCE | | |
|--|---|--|
| COLUMN INFORMATION SERIAL NUMBER : 2002000204 PART NUMBER : C1515-B5032 COLUMN TYPE : ChraSil B100 C8, 150 x 4.6 mm, 5.0um LOT : KL0051 | | TEST CONDITIONS MOBILE PHASE : COLUMN FLOW RATE : COLUMN FLOW RATE : LINEAR VELOCITY : TEMPERATURE : INJECTION VOLUME : |
| TEST VALUES THEORETICAL PLATES : 13800 SELECTIVITY : 3.27 USP TAILING FACTOR : 1.05 (@ 5% Peak Height) | SPECIFICATIONS MIN : 91998 RANGE : 3.02 - 3.62 RANGE : 0.98 - 1.25 k' : 3.39 | Peak 1. 2. 3. 4. |
|  Image not found or type unknown | | |
| MATERIAL SAFETY DATA SHEETS ARE AVAILABLE | | |
| ChraReport at https://chracert.com/2002000204 | | |

Revision #4

Created 9 August 2021 13:23:52 by Admin

Updated 6 October 2021 04:25:31 by Admin