


Product Specification Sheet

by ChraWork a Gimium™ Ecosystem

2002000214

LC COLUMN QUALITY CONTROL PERFORMANCE		
COLUMN INFORMATION SERIAL NUMBER : 2002000214 PART NUMBER : C1515-B5032 COLUMN TYPE : ChraSil B100 C8, 150 x 4.6 mm, 5.0um LOT : KL0243		TEST CONDITIONS MOBILE PHASE : COLUMN FLOW RATE : COLUMN FLOW RATE : LINEAR VELOCITY : TEMPERATURE : INJECTION VOLUME :
TEST VALUES THEORETICAL PLATES : 13300 SELECTIVITY : 3.27 USP TAILING FACTOR : 1.05 (@ 5% Peak Height)	SPECIFICATIONS MIN : 91398 RANGE : 3.02 - 3.62 RANGE : 0.98 - 1.25 k' : 3.22	Peak 1. 2. 3. 4.
 Image not found or type unknown		
MATERIAL SAFETY DATA SHEETS ARE AVAILABLE		
ChraReport at https://chracert.com/2002000214		

Revision #5

Created 9 August 2021 13:18:14 by Admin

Updated 6 October 2021 04:25:15 by Admin