


Product Specification Sheet

by ChraWork a Gimium™ Ecosystem

2012000304

LC COLUMN QUALITY CONTROL PERFORMANCE		
COLUMN INFORMATION SERIAL NUMBER : 2012000304 PART NUMBER : C1505-B5032 COLUMN TYPE : ChraSil B100 C18, 150 x 4.6 mm, 5.0um LOT : VD184003		TEST CONDITIONS MOBILE PHASE : COLUMN FLOW RATE : COLUMN FLOW RATE : LINEAR VELOCITY : TEMPERATURE : INJECTION VOLUME :
TEST VALUES THEORETICAL PLATES : 17455 SELECTIVITY : 3.27 USP TAILING FACTOR : 1.13 (@ 5% Peak Height)	SPECIFICATIONS MIN : 16780 RANGE : 3.02 - 3.62 RANGE : 0.98 - 1.25 k' : 2.89	Peak 1. 2. 3. 4.
 Image not found or type unknown		
MATERIAL SAFETY DATA SHEETS ARE AVAILABLE		
ChraReport at https://chracert.com/2012000304		

Revision #18

Created 9 August 2021 07:55:12 by Admin

Updated 6 October 2021 04:24:14 by Admin