


2012000324

LC COLUMN QUALITY CONTROL PERFORMANCE		
COLUMN INFORMATION SERIAL NUMBER : 2012000324 PART NUMBER : C1505-B5034 COLUMN TYPE : ChraSil B100 C18, 250 x 4.6 mm, 5.0um LOT : KL146923		TEST CONDITIONS MOBILE PHASE : COLUMN FLOW RATE : COLUMN FLOW RATE : LINEAR VELOCITY : TEMPERATURE : INJECTION VOLUME :
TEST VALUES THEORETICAL PLATES : 13800 SELECTIVITY : 3.27 USP TAILING FACTOR : 1.05 (@ 5% Peak Height)	SPECIFICATIONS MIN : 91998 RANGE : 3.02 - 3.62 RANGE : 0.98 - 1.25 k' : 3.39	Peak 1. 2. 3. 4.
 Image not found or type unknown		
MATERIAL SAFETY DATA SHEETS ARE AVAILABLE		
ChraReport at https://chracert.com/2012000324		

Revision #5

Created 9 August 2021 13:14:39 by Admin

Updated 6 October 2021 04:24:53 by Admin