

**Product Specification Sheet**

by ChraWork a Gimium™ Ecosystem

# 2207000362

LC COLUMN QUALITY CONTROL PERFORMANCE		
<b>COLUMN INFORMATION</b> SERIAL NUMBER : 2207000362 PART NUMBER : C1505-B5032 COLUMN TYPE : ChraSil B100 C18, 150 x 4.6 mm, 5.0um LOT : KL0051		<b>TEST CONDITIONS</b> MOBILE PHASE : 0.1% TFA / 99.9% H <sub>2</sub> O COLUMN FLOW RATE : 1.0 mL/min COLUMN FLOW RATE : 1.0 mL/min LINEAR VELOCITY : 0.5 cm/min TEMPERATURE : 30°C INJECTION VOLUME : 10 µL
<b>TEST VALUES</b> THEORETICAL PLATES : 109277 SELECTIVITY : 2.17 USP TAILING FACTOR : 1.05 (@ 5% Peak Height)	<b>SPECIFICATIONS</b> MIN : 16392 RANGE : 3.02 - 3.62 RANGE : 0.98 - 1.25 k' : 2.22	<b>Peak</b> 1. 2. 3. 4.
 Image not found or type unknown		
MATERIAL SAFETY DATA SHEETS ARE AVAILABLE		
<a href="https://chracert.com/2207000362">ChraCert</a> at <a href="https://chracert.com/2207000362">https://chracert.com/2207000362</a>		

Revision #2

Created 26 July 2022 01:47:17 by Admin

Updated 26 July 2022 01:50:48 by Admin