

C0715-D5006

| | |
|--------------------|--|
| Description | ChraSil HPLC Colum |
| Manufacturer | Chratec |
| Internal Diameter | 3.0 |
| Column Length | 100 |
| Particle Size | 5.0 |
| Modification | C8-3 |
| Pressure Stability | 400 bar |
| Base Material | Silica, spherical, poi |
| Packing brand | ChraSil |
| USP L Code | L7 |
| Part Number | C0715-D5006 |
| pH Range# | 2 - 8 |
| Carbon Load# | - (See https://chratec.com) |
| Endcap | # |
| Pore Size | - |
| Surface Area(m2/g) | - |
| Column Group# | - |
| ChraSil URL# | C0715-D5006 at ht |
| Installation | Please see instrume |
| Replacement | Beyond routine mai |

| | |
|--------------|---|
| Care and Use | https://chrates.com |
| Warranty# | See ChraContact at |

Revision #1

Created 30 July 2021 06:36:26 by Admin

Updated 30 July 2021 06:36:35 by Admin