

# C1195-C6003

|                     |                                |
|---------------------|--------------------------------|
| Description         | ChraSil C120 C18, 150 x 2.1 mm |
| Product Type        | HPLC Column (Liquid Chromatog  |
| For Use with        | HPLC Instrument                |
| Manufacturer        | Chratec                        |
| Internal Diameter   | 2.1                            |
| Column Length       | 150                            |
| Particle Size       | 5.0                            |
| Modification        | C8                             |
| Pressure Stability  | 400 bar                        |
| Base Material       | Silica, spherical, porous      |
| Packing brand       | ChraSil                        |
| USP L Code          | L1                             |
| pH Range#           | 2 - 8                          |
| Carbon Load#        | 11.0 %                         |
| Endcap              | + (Yes)                        |
| Pore Size           | 130Å                           |
| Surface Area(m2/g)  | 170                            |
| Operating Temp (°C) | 20 - 70                        |
| Column Group#       | C120                           |
| Part Number         | C1195-C6003                    |

|              |   |
|--------------|---|
| Installation | Please see instrument manual for  |
| Replacement  | Beyond routine maintenance  |
| Care and Use | <a href="https://chratesc.com/mag/CaU.pdf">https://chratesc.com/mag/CaU.pdf</a> |
| Warranty#    | See ChraContact at <a href="https://chratesc.com">https://chratesc.com</a>      |
| ChraSil URL# | C1195-C6003 at <a href="https://chratesc.com">https://chratesc.com</a>          |

Revision #1

Created 11 November 2021 02:09:50 by Admin

Updated 11 November 2021 02:10:03 by Admin